

# Detailed Test History



Test Status: Fail, Test Type: All  
Person: All, Instrument: All  
From: 24/06/2011 To: 24/06/2011

<b>Client:</b>	<b>Site: BLOCK A</b>	<b>Location: LABORATORY</b>	
Asset ID: MI070	Description: PENETROMETER	User: OLIVER	Test Instrument: X06-1324
Comments: HUMBOLDT PENETROMETER HAS POWER CORD UNCLAMPED AT PLUG. LEFT IN SERVICE. RECOMMEND TO REPAIR.			

Test Type	Date	Result	Unit	Status
Visual	24/06/2011			Fail
Overall: Fail				

Asset ID: MI071	Description: IMPACT TESTER	User: OLIVER	Test Instrument: X06-1324
Comments: LABTHINK IMPACT TESTER HAS POWER CORD UNCLAMPED AT PLUG. LEFT IN SERVICE. RECOMMEND TO REPAIR.			

Test Type	Date	Result	Unit	Status
Visual	24/06/2011			Fail
Overall: Fail				

Asset ID: MI074	Description: IEC LEAD	User: OLIVER	Test Instrument: X06-1324
Comments: OVEN LEAD HAS NO EARTH CONNECTION. REMOVED FROM SERVICE. RECOMMEND TO REPLACE.			

Test Type	Date	Result	Unit	Status
Visual	24/06/2011			Pass
Earth Current	24/06/2011	25	A	Information
Earth	24/06/2011	>19.99	Ohm	Fail
Overall: Fail				

Asset ID: MI075	Description: IEC LEAD	User: OLIVER	Test Instrument: X06-1324
Comments: OVEN LEAD IS DAMAGED. REMOVED FROM SERVICE. RECOMMEND TO REPLACE.			

Test Type	Date	Result	Unit	Status
Visual	24/06/2011			Fail
Overall: Fail				

Asset ID: MI077	Description: OVEN	User: OLIVER	Test Instrument: X06-1324
Comments: THERMO OVEN No.07 FAILS INSULATION TEST. POSSIBLE HEATING ELEMENT FAILURE. RECOMMEND TO REPAIR.			

Test Type	Date	Result	Unit	Status
Visual	24/06/2011			Pass
Earth Current	24/06/2011	25	A	Information
Earth	24/06/2011	0.09	Ohm	Pass
Insulation	24/06/2011	0.37	MOhm	Fail
Overall: Fail				

Asset ID: MI079	Description: IEC LEAD	User: OLIVER	Test Instrument: X06-1324
Comments: OVEN LEAD IS DAMAGED. REMOVED FROM SERVICE. RECOMMEND TO REPLACE.			

Test Type	Date	Result	Unit	Status
Visual	24/06/2011			Fail
Overall: Fail				

Asset ID: MI080	Description: IEC LEAD	User: OLIVER	Test Instrument: X06-1324
Comments: OVEN LEAD IS DAMAGED. REMOVED FROM SERVICE. RECOMMEND TO REPLACE.			

Test Type	Date	Result	Unit	Status
Visual	24/06/2011			Fail
Overall: Fail				

Total Tests: 7